



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

STANDRIDGE GRANITE CORP.  
9437 Santa Fe Springs Road  
Santa Fe Springs, CA 90670  
Flor Espinoza-Carrales Phone: 562 946 6334

CALIBRATION

Valid To: March 31, 2026

Certificate Number: 3559.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 6</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Granite Surface Plates <sup>3</sup> –			
Flatness	Up to 480 in	25 µin	Autocollimator
Repeat Reading*	Up to 480 in	19 µin	Repeat-o-meter, *Only valid in connection with flatness calibration

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
Granite Accessories –			
Straightness	Up to 144 in	16 $\mu$ in	Autocollimator
Perpendicularity	Up to 144 in	30 $\mu$ in	Autocollimator with penta prism on surface plate, $L$ in inches
	Up to 6 in	120 $\mu$ in	Squarol & Master V block
Parallelism	Up to 48 in	31 $\mu$ in	Electronic amplifier
	Up to 168 in	32 $\mu$ in	Autocollimator
Length	Up to 12 in	(53 + 0.20 $L$ ) $\mu$ in	Gage blocks

<sup>1</sup> This laboratory offers commercial and field calibration services and new manufactured products.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately a 95 % level of confidence, usually using a coverage factor of  $k=2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the uncertainties achievable on a customer's site can normally be expected to be larger than the CMC that the accredited laboratory has been assigned on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the CMC.

<sup>4</sup> In the statement of CMC, the numerical value  $L$  is for the nominal length of the device measured in inches.

<sup>5</sup> The CMC values listed do NOT include long-term stability of the UUT, e.g., for surface plates the tolerances are not included.

<sup>6</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



## Accredited Laboratory

A2LA has accredited

### STANDRIDGE GRANITE CORP.

*Santa Fe Springs, CA*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 26<sup>th</sup> day of March 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 3559.01  
Valid to March 31, 2026

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*