



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

FAIRBANKS SCALES, INC.
490 Bullis Road
West Seneca, NY 14224
Joe Ehrnschwender Phone: 513 860 8062

CALIBRATION

Valid To: December 31, 2019

Certificate Number: 1843.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Balances ³	(0 to 100) g	0.30 mg	Class 1, 6 and F weights
	(100 to 500) g	0.10 g	
	(500 to 4000) g	0.11 g	
	(4000 to 25 000) g	3.0 g	
Scales ³	(0 to 100) lb	0.011 lb	Class 6 and F weights
	(101 to 500) lb	0.035 lb	Class F weights and substitution
	(501 to 1000) lb	0.12 lb	
	(1001 to 5000) lb	0.39 lb	
	(5001 to 10 000) lb	0.69 lb	
	(10 001 to 200 000) lb	50 lb	

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

WITHDRAWN

A handwritten signature in black ink, appearing to be 'L. L. L.', located in the bottom right corner of the page.



Accredited Laboratory

A2LA has accredited

FAIRBANKS SCALES, INC.

West Seneca, NY

for technical competence in the field of

Calibration
WITHDRAWN

This laboratory is accredited in accordance with the recognized international Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 27th day of December 2017.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 1843.01
Valid to December 31, 2019

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.