



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

FIRSTENERGY BETA LABORATORY  
6670 Beta Drive  
Mayfield Village, OH 44143  
Timothy Gervan Phone: 440 604 9729

CALIBRATION

Valid To: November 30, 2020

Certificate Number: 2489.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,5</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Micrometers – Inside Outside	Up to 12 in Up to 12 in	(300 + 150L + 0.60R) μin (340 + 44L + 0.60R) μin	Comparison to gage blocks
Calipers	Up to 12 in	(340 + 44L + 0.60R) μin	Comparison to gage blocks
Dial Indicators	Up to 2 in	(90 + 14L + 0.60R) μin	Comparison to gage blocks
Depth Gages	Up to 12 in	(430 + 83L + 0.60R) μin	Comparison to gage blocks

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
DC Voltage – Generate	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	19 μV/V + 0.50 μV 12 μV/V + 1.2 μV 4.8 μV/V + 5.1 μV 9.4 μV/V + 6.7 μV 14 μV/V + 48 μV 7.6 μV/V + 1.2 mV	Fluke 5720A
DC Current – Generate	(0 to 220) μA (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	46 μA/A + 8.0 nA 51 μA/A + 59 nA 50 μA/A + 0.12 μA 70 μA/A + 1.4 μA 0.018 % + 23 μA	Fluke 5720A
DC Voltage – Measure	(0 to 200) mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1000) V	13 μV/V + 0.20 μV 6.1 μV/V + 0.50 μV 5.7 μV/V + 5.0 μV 8.6 μV/V + 43 μV 7.0 μV/V + 0.53 mV	Fluke 8508A
DC Current – Measure	(0 to 200) μA (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A (2 to 20) A	23 μA/A + 1.0 nA 23 μA/A + 5.0 nA 29 μA/A + 41 nA 69 μA/A + 0.90 μA 0.032 % + 17 μA 0.071 % + 0.50 mA	Fluke 8508A
Capacitance – Generate	(0.33 to 0.4999) nF (0.5 to 1.0999) nF (1.1 to 3.2999) nF (3.3 to 10.999) nF (11 to 32.999) nF (33 to 109.99) nF (110 to 329.99) nF (0.33 to 1.0999) μF (1.1 to 3.2999) μF (3.3 to 10.999) μF (11 to 32.999) μF (33 to 109.99) μF (110 to 329.99) μF (0.33 to 1.1) mF	0.57 % + 0.020 nF 1.1 % + 0.020 nF 0.86 % + 0.020 nF 0.69 % + 0.020 nF 0.65 % + 0.20 nF 0.40 % + 0.20 nF 0.40 % + 0.40 nF 0.40 % + 2.0 nF 0.57 % + 4.0 nF 0.52 % + 12 nF 0.60 % + 36 nF 0.75 % + 120 nF 1.1 % + 380 nF 1.4 % + 570 nF	Fluke 5520A

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Resistance – Measure	(0 to 2) Ω (2 to 20) Ω (20 to 200) Ω 200 Ω to 2 kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	37 μΩ 0.21 mΩ 1.7 mΩ 18 mΩ 0.17 Ω 1.7 Ω 21 Ω 430 Ω 23 kΩ 2.1 MΩ	Fluke 8508A
Resistance – Generate	<p>Fixed Points</p> 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	1.3 Ω 2.5 Ω 24 Ω 48 Ω 0.50 kΩ 1.1 kΩ 12 kΩ	Fluke 5720A
4 Wire	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ	74 μΩ 0.11 mΩ 0.21 mΩ 0.27 mΩ 0.51 mΩ 1.2 mΩ 2.3 mΩ 10 mΩ 19 mΩ 0.10 Ω 0.20 Ω	
Temperature – RTD Simulated Generate	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.050 °C 0.067 °C 0.077 °C 0.089 °C 0.12 °C 0.14 °C 0.17 °C	ESI RS925D (based on platinum RTD 385)

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Electrical Calibration of Thermocouple Indication Systems –			
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.52 °C 0.40 °C 0.35 °C 0.39 °C	Fluke 5520A
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.58 °C 0.19 °C 0.17 °C 0.19 °C 0.25 °C	
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.32 °C 0.19 °C 0.17 °C 0.21 °C 0.27 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.39 °C 0.22 °C 0.19 °C 0.31 °C 0.47 °C	
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.47 °C 0.26 °C 0.23 °C 0.22 °C 0.32 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.66 °C 0.41 °C 0.39 °C 0.47 °C	
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.55 °C 0.42 °C 0.43 °C 0.54 °C	
Type T	(-250 to 150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.73 °C 0.28 °C 0.19 °C 0.17 °C	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Measure			
(0 to 200) mV	10 Hz to 30 kHz (30 to 100) kHz	0.040 % + 12 μV 0.10 % + 26 μV	Fluke 8508A
(0.2 to 2) V	10 Hz to 30 kHz (30 to 100) kHz 100 kHz to 1 MHz	0.024 % + 20 μV 0.060 % + 0.30 mV 1.8 % + 27 mV	
(20 to 20) V	10 Hz to 30 kHz (30 to 100) kHz 100 kHz to 1 MHz	0.23 % + 270 μV 0.060 % + 3.0 mV 1.9 % + 0.30 V	
(20 to 200) V	10 Hz to 30 kHz (30 to 100) kHz	0.024 % + 5.3 mV 0.060 % + 30 mV	
(200 to 1000) V	10 Hz to 10 kHz (10 to 30) kHz	0.050 % + 27 mV 0.30 % + 27 mV	
AC Voltage – Generate			
(0 to 2.2) mV	10 Hz to 100 kHz 100 kHz to 1 MHz	0.47 % + 6.0 μV 2.6 % + 25 μV	Fluke 5720A
(2.2 to 22) mV	10 Hz to 100 kHz 100 kHz to 1 MHz	0.050 % + 7.0 μV 0.28 % + 24 μV	
(22 to 220) mV	10 Hz to 100 kHz 100 kHz to 1 MHz	0.047 % + 21 μV 0.28 % + 68 μV	
(0.22 to 2.2) V	10 Hz to 100 kHz 100 kHz to 1 MHz	0.029 % + 39 μV 0.20 % + 520 μV	
(2.2 to 22) V	10 Hz to 100 kHz 100 kHz to 1 MHz	0.030 % + 290 μV 0.17 % + 6.4 mV	
(22 to 220) V	10 Hz to 100 kHz	0.018 % + 4.0 mV	
(220 to 1100) V	15 Hz to 1 kHz	0.023 % + 13 mV	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
<b>AC Current – Measure</b>  (0 to 200) µA (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A (2 to 20) A	10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 10 kHz	0.042 % + 48 nA 0.052 % + 210 nA 0.045 % + 3.1 µA 0.043 % + 27 µA 1.7 mA/A + 240 µA 4.4 mA/A + 3.0 mA	Fluke 8508A
<b>AC Current – Generate</b>  (9 to 220) µA  (0.22 to 2.2) mA  (2.2 to 22) mA  (22 to 220) mA  (0.22 to 2.2) A	10 Hz to 5 kHz (5 to 10) kHz  10 Hz to 5 kHz (5 to 10) kHz  10 Hz to 5 kHz (5 to 10) kHz  10 Hz to 5 kHz (5 to 10) kHz  20 Hz to 5 kHz (5 to 10) kHz	0.070 % + 24 nA 0.16 % + 0.10 µA  0.024 % + 200 nA 0.14 % + 1.0 µA  0.022 % + 1.7 µA 0.13 % + 8.6 µA  0.022 % + 15 µA 0.13 % + 45 µA  0.048 % + 250 µA 0.83 % + 460 µA	Comparison to a multifunction calibrator Fluke 5720A

III. Fluid Quantities

Parameter/Equipment	Range	CMC <sup>2,6</sup> (±)	Comments
Flow – Measure			
Air	(1 to 20) sccm	0.050 sccm	DH Molbloc system
	(20 to 50) sccm	0.12 sccm	
	(50 to 100) sccm	0.23 sccm	
	(100 to 150) sccm	0.35 sccm	
	(150 to 200) sccm	0.46 sccm	
	(0.2 to 0.5) slm	0.0020 slm	
	(0.5 to 1) slm	0.0030 slm	
	(1 to 1.5) slm	0.0040 slm	
	(1.5 to 2) slm	0.0050 slm	
	(2 to 5) slm	0.060 slm	
	(5 to 10) slm	0.11 slm	
	(10 to 15) slm	0.16 slm	
	(15 to 20) slm	0.22 slm	
	(20 to 25) slm	0.070 slm	
	(25 to 30) slm	0.090 slm	
	(30 to 35) slm	0.10 slm	
	(35 to 40) slm	0.12 slm	
	(40 to 50) slm	0.12 slm	
	(50 to 60) slm	0.14 slm	
	(60 to 70) slm	0.16 slm	
	(70 to 80) slm	0.18 slm	
	(80 to 90) slm	0.21 slm	
	(90 to 100) slm	0.23 slm	
	(100 to 125) slm	0.30 slm	
	(125 to 150) slm	0.36 slm	
	(150 to 175) slm	0.42 slm	
	(175 to 200) slm	0.48 slm	
	(200 to 225) slm	0.54 slm	
	(225 to 250) slm	0.59 slm	
	(250 to 300) slm	0.71 slm	
	(300 to 400) slm	0.95 slm	
	(400 to 500) slm	1.2 slm	
	(500 to 600) slm	1.5 slm	
	(600 to 700) slm	1.7 slm	
	(700 to 800) slm	1.9 slm	
	(800 to 900) slm	2.2 slm	
	(900 to 1000) slm	2.4 slm	

IV. Mechanical

Parameter/Equipment	Range	CMC <sup>2,6</sup> (±)	Comments
Vacuum – Measure and Measuring Equipment	Up to 1000 µm Hg	1.0 % of reading	MKS 660B/627B direct reading vacuum gage
	(1 to 10) mm Hg	0.50 % of reading	MKS 660B/626A direct reading vacuum gage
Scales and Balances – Fixed Values (0 to 20) kg*	(1 to 500) mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg	0.14 mg 0.16 mg 0.19 mg 0.24 mg 0.29 mg 0.38 mg 0.63 mg 1.1 mg 2.1 mg 7.8 mg 15 mg 23 mg 53 mg	Class 3 weights *fixed values uncertainties are per weight
(0 to 2500) lb*	0.5 lb 1 lb 2 lb 5 lb 10 lb 20 lb 50 lb	0.000 11 lb 0.000 16 lb 0.000 21 lb 0.000 52 lb 0.0011 lb 0.0022 lb 0.0053 lb	Class F weights *fixed values uncertainties are per weight
Pressure – Measure and Measuring Equipment  Pneumatic	(2.5 to 9.5) psig (9.5 to 1023) psig	80 parts in 10 <sup>6</sup> psig 45 parts in 10 <sup>6</sup> psig	2465 piston gauge



Parameter/Equipment	Range	CMC <sup>2,6</sup> (±)	Comments
Pressure – Measure and Measuring Equipment (cont)			
Hydraulic Only	(1000 to 22 000) psig	45 parts in 10 <sup>6</sup> psig	DHI PG7302
Torque – Measuring Equipment			
Counter Clockwise and Clockwise	(1.25 to 4) in·lbf (4 to 8) in·lbf (8 to 12) in·lbf (12 to 16) in·lbf (16 to 20) in·lbf  (20 to 40) in·lbf (40 to 60) in·lbf (60 to 80) in·lbf (80 to 100) in·lbf (100 to 120) in·lbf  (10 to 20) ft·lbf (20 to 40) ft·lbf (40 to 60) ft·lbf (60 to 80) ft·lbf (80 to 100) ft·lbf  (100 to 120) ft·lbf (120 to 180) ft·lbf (180 to 240) ft·lbf (240 to 300) ft·lbf  (300 to 400) ft·lbf (400 to 600) ft·lbf (600 to 800) ft·lbf (800 to 1000) ft·lbf  (1000 to 1600) ft·lbf (1600 to 2400) ft·lbf (2400 to 3200) ft·lbf (3200 to 4000) ft·lbf  (4000 to 5200) ft·lbf (5200 to 6500) ft·lbf	0.10 in·lbf 0.10 in·lbf 0.13 in·lbf 0.14 in·lbf 0.16 in·lbf  0.21 in·lbf 0.49 in·lbf 0.65 in·lbf 0.81 in·lbf 0.97 in·lbf  0.16 ft·lbf 0.32 ft·lbf 0.49 ft·lbf 0.65 ft·lbf 0.81 ft·lbf  0.97 ft·lbf 1.5 ft·lbf 2.0 ft·lbf 2.5 ft·lbf  3.3 ft·lbf 4.9 ft·lbf 6.5 ft·lbf 8.1 ft·lbf  13 ft·lbf 20 ft·lbf 26 ft·lbf 33 ft·lbf  42 ft·lbf 53 ft·lbf	Torque arms and class F weights

Parameter/Equipment	Range	CMC <sup>2,6</sup> (±)	Comments
Torque – Measuring Equipment (cont)  Clockwise Only	(6500 to 8000) ft·lbf (8000 to 12 000) ft·lbf (12 000 to 16 000) ft·lbf (16 000 to 20 000) ft·lbf	95 ft·lbf 150 ft·lbf 190 ft·lbf 240 ft·lbf	Torque arms and class F weights
Torque Wrenches and Hand Tools	(1.25 to 120) in·lbf (10 to 300) ft·lbf (300 to 1000) ft·lbf (1000 to 4000) ft·lbf	2.8 in·lbf 2.8 ft·lbf 8.2 ft·lbf 78 ft·lbf	AKO torque transducers and test stand

#### V. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2,6</sup> (±)	Comments
Temperature – Measure	(-20 to 0) °C (0 to 300) °C (300 to 500) °C (500 to 550) °C	0.030 °C 0.040 °C 0.040 °C 0.040 °C	Circulating bath/SPRT

#### VI. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Frequency – Measuring Equipment	1 Hz to 80 MHz  10 MHz	4.4 mHz  6.8 pHz/Hz	Fluke 910R/ synthesizer
Frequency – Measure	1 Hz to 100 kHz 100 kHz to 1 MHz (1 to 100) MHz (100 to 225) MHz	14 μHz 150 μHz 6.8 mHz 17 mHz	Fluke 910R /counter

---

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in inches and  $R$  is the value of the resolution of the device in microinches.

<sup>4</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC are expressed as either a specific value that covers the full range or as a fraction/percentage of the reading plus a fixed floor specification.

<sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.

<sup>6</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



## Accredited Laboratory

A2LA has accredited

### FIRSTENERGY BETA LABORATORY

*Mayfield Village, OH*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 21<sup>st</sup> day of December 2018.

A blue ink signature of the Senior Director of Accreditation Services.

Senior Director, Accreditation Services  
For the Accreditation Council  
Certificate Number 2489.01  
Valid to November 30, 2020

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*